

AperTO - Archivio Istituzionale Open Access dell'Università di Torino

## A simple method to produce pectin-honey hydrogels and its characterization as new biomaterial for surgical use

### This is the author's manuscript

*Original Citation:*

*Availability:*

This version is available <http://hdl.handle.net/2318/1570083> since 2016-06-22T11:13:52Z

*Published version:*

DOI:10. 1159/446131

*Terms of use:*

Open Access

Anyone can freely access the full text of works made available as "Open Access". Works made available under a Creative Commons license can be used according to the terms and conditions of said license. Use of all other works requires consent of the right holder (author or publisher) if not exempted from copyright protection by the applicable law.

(Article begins on next page)

# **A simple method to produce pectin-honey hydrogels and its characterization as new biomaterial for surgical use**

Gessica Giusto<sup>a</sup>, Giangiacomo Beretta<sup>b</sup>, Cristina Vercelli<sup>a</sup>, Emanuela Valle<sup>a</sup>, Roberta Borghi<sup>c</sup>, Patrizio Odetti<sup>c</sup>, Fiammetta Monacelli<sup>c</sup>, Clara Tramuta<sup>a</sup>, Elena Grego<sup>a</sup>, Patrizia Nebbia<sup>a</sup>, Patrizia Robino<sup>a</sup>, Marco Gandini<sup>a</sup>

**Keywords:** Honey-based material, Biocompatibility, Manuka honey,

**Background:** Novel honey-based membranes have been developed and characterized as a medical device. In this study, a novel, simple and fast method to produce pectin-honey wound dressings is described. An ideal wound dressing is yet to be developed. Ideally, a wound dressing should maintain optimal fluid affinity, permit moisture evaporation, protect the wound from foreign microbes, and have shape-conformability, biocompatibility, and antibacterial activity.

**Materials and Method:** The properties of these honey-based membranes were systematically investigated with respect to physical and chemical parameters, including swelling ability, water vapour transmission rate, hydrogen peroxide production, methylglyoxal content and antibacterial activity. Furthermore, these membranes were assessed for cytocompatibility by performing a proliferation study using cultured fibroblast cells.

**Results:** The new membranes may be used as wound dressings as they have a good WVTR and fluid uptake and have no proven cytotoxicity to fibroblasts. The membranes demonstrate good antibacterial activity toward clinically relevant pathogenic microorganisms such as *S. aureus* and *E. coli*.

**Conclusion:** In conclusion, the honey-based membranes can be applied as a medical device for fabrication of wound dressings. The presence of a natural active component, conformability, and complete resorbability are the characteristics of this new biocompatible biomaterial that respects the pathophysiology of wounds, improves healing, and does not cause pain upon removal. Furthermore, the production of these device is extremely simple and inexpensive.